

DIN 58345:2007

DECLARATION OF CONFORMITY

Document no.: DK221006
Manufacturer: A/S Vestfrost
Address: Falkevej 12
6705 Esbjerg Ø
Denmark
Product designation: See Enclosure No. 1.

A/S Vestfrost declare under sole responsibility that Pharmacy Refrigerators comply with DIN 58345.

DIN stands for "Deutsches Institut für Normung" (German Institute for Standardization). The Institute's specialist bodies define industry standards. DIN 58345 relates to pharmaceutical refrigerators that are used to store pharmaceuticals at an operating temperature of 2°C to 8°C. DIN 58345 currently reflects the latest technological refrigerators.

Basic requirements established by DIN 58345:

- Operating temperature: +2°C to +8°C
- Usable in ambient temperature: +10°C to +35°C
- Noise emissions: < 60dB(A)
- Visual and acoustic warning system for high and low temperatures
- Dry contact for external alarm systems
- Warning system active from min. 12 h in the event of a power cut
- Safety thermostat for preventing temperatures below +2°C
- Load bearing capacity at least 100 kg/m²
- Lockable door

A/S VESTFROST

2022.10.06

Appliance Certification Manager

Jørgen Dyrholm

name

Esbjerg Denmark

VESTFROST
SOLUTIONS

A/S Vestfrost
Falkevej 12

DK-6705 Esbjerg Ø, Denmark

Tel.: +45 79 14 22 22 · Fax: +45 79 14 23 55

signature

Esbjerg Denmark

Enclosure No. 1. Declaration of Conformity Date 06-10-2022

<u>Identification</u>	<u>Type or product No.</u>	<u>Trade name.</u>
VESTFROST	AKS 070i	R-90i DIN
VESTFROST	AKS 110i	R-138i DIN
VESTFROST	AKS 210i	R-427i DIN